

Atrial Arrhythmias In COVID-19 Patients And The Value Of Ultrasound

The articles in this email may consist of clinical trials and observational studies as well as economic analyses. Access to these articles and to additional clinical articles with similar results are provided through the embedded links. Membership to individual sites may be required in order to access the material.

Atrial Arrhythmias in COVID-19 Patients

Colon CM, Barrios JG, Chiles JW, et al. [Atrial arrhythmias in COVID-19 Patients](#). *JACC Clinical Electrophysiol.* 2020; In press. doi: 10.1016/j.jacep.2020.05.015.



OBJECTIVE

- Analyze electrocardiographs (ECG) and telemetry to better understand the occurrence and impact of atrial arrhythmias in patients with the SARS-CoV-2 (COVID-19) infection



METHOD

- Observational analysis of 115 patients with COVID-19 admitted to the University of Alabama Birmingham Hospital (69 to intensive care unit, 46 to general ward)



RESULTS



16.5%

of all admitted patients developed atrial tachyarrhythmias during their hospital stay (100% of those admitted to intensive care unit vs. 0% of those admitted to general ward)

Of the atrial tachyarrhythmia patients:

63.1% had atrial fibrillations

79.0% needed vasopressors to regulate blood pressure

52.6% became hemodynamically compromised

26.3% ultimately died

Strong association existed between developing atrial arrhythmias and needing mechanical ventilation ($P=0.0002$)



KEY TAKEAWAYS

- Atrial arrhythmias are common in COVID-19 patients admitted to the intensive care unit and not common in COVID-19 patients admitted to the general ward.
- Atrial arrhythmias in severely ill COVID-19 patients were associated with additional serious complications.

ARTICLES WITH SIMILAR RESULTS

Inciardi R, Adamo M, Lupi L, et al. [Characteristics and outcomes of patients hospitalized for COVID-19 and cardiac disease in Northern Italy](#). *European Heart Journal*. 2020;41(19):1821-1829. doi:10.1093/eurheartj/ehaa388.



MEDICAL SOCIETY GUIDANCE ON CARDIAC IMAGING DURING COVID-19

- A recent statement by the **American Society of Echocardiography (ASE)** has advised that transesophageal echocardiograms (TEE) increase the risk of exposure to COVID-19.
- A growing number of **medical societies** (including the European Society of Cardiology, the British Society of Echocardiography, the Canadian Society of Echocardiography, and the Cardiac Society of Australia and New Zealand) **are cautioning against the use of TEE and recommending alternative imaging modalities such as ultrasound for cardiac procedures.**
- Read the ASE statement here: <https://www.asecho.org/ase-statement-covid-19/>.



FORWARD TO A FRIEND

The Use of Intracardiac Echocardiography (ICE) Catheters in Endocardial Ablation of Cardiac Arrhythmia: Meta-Analysis of Efficiency, Effectiveness, and Safety Outcomes

Goya M, Frame D, Gache L, et al. [The use of intracardiac echocardiography catheters in endocardial ablation of cardiac arrhythmia: Meta-analysis of efficiency, effectiveness, and safety outcomes](#). *J Cardiovasc Electrophysiol*. 2020;31(3):664-673. doi:10.1111/jce.14367.



OBJECTIVE

- Assess procedural benefits associated with the use of ICE during catheter ablation of cardiac arrhythmias compared to ablation without the use of ICE



METHOD

- Systematic literature review and meta-analysis
- Identified published studies from 1996-2018 via PubMed/MEDLINE
- Study eligibility: comparative studies that assessed the use of ICE during the catheter ablation of cardiac arrhythmias vs. ablation without ICE use



RESULTS

Use of ICE is associated with:

Reduction in fluoroscopy time



An average **6.95 minute reduction** in fluoroscopy time ($P<0.01$)

12.7 minute reduction in fluoroscopy time when **sensor-based ICE catheters**, such as the SOUNDSTAR® Catheter, are used ($P=0.02$)

Reduction in procedure time



An average **15.2 minute reduction** in procedure time ($P<0.01$)



A trend of decreased complications was noted with the use of ICE; however, this difference did not reach statistical significance ($P=0.08$).



KEY TAKEAWAYS

- The use of ICE during ablation of cardiac arrhythmias is associated with **lower fluoroscopy time and shorter procedure time** without compromising safety.

ARTICLES WITH SIMILAR RESULTS

Lyan E, Tsyganov A, Abdrahmanov A, et al. [Nonfluoroscopic Catheter Ablation of Paroxysmal Atrial Fibrillation](#). *Pacing Clin Electrophysiol*. 2018;41(6):611-619. doi:10.1111/pace.13321.



FORWARD TO A FRIEND

In the US, 4mm Catheters (NAVISTAR® Catheter, CELSIUS® Catheter, EZ STEER® Catheter (NAV and Non-NAV)) have a “General Indication” for creation of endocardial lesions in patients 4 years of age and older. This “General Indication” includes treatment of Ventricular Tachycardia.

THERMOCOOL® Navigation Catheters are indicated for the treatment of drug refractory recurrent symptomatic paroxysmal atrial fibrillation, when used with CARTO® 3 Systems (excluding NAVISTAR® RMT THERMOCOOL® Catheter).

Bertaglia E, Bella PD, Tondo C, Proclemer A, Bottoni N, De Ponti R, et al. Image integration increases efficacy of paroxysmal atrial fibrillation catheter ablation: results from the CARTOMERGE® Module Italian Registry. *Europace* 2009;11:1004-1010.

THERMOCOOL® Navigation Catheters are indicated for the treatment of recurrent drug/device refractory sustained monomorphic ventricular tachycardia (VT) due to prior myocardial infarction (MI) in adults

The NAVISTAR® THERMOCOOL® and EZ STEER® THERMOCOOL® NAV Catheters are FDA approved for the treatment of drug refractory recurrent symptomatic paroxysmal atrial fibrillation, when used with compatible three-dimensional electroanatomic mapping systems.

In the US, Biosense Webster THERMOCOOL® Non-Navigational Catheters are indicated for the treatment of Type I Atrial Flutter in patients 18 years of age or older.

Always verify catheter tip location using fluoroscopy or IC signals and consult the CARTO® 3 System User Guide regarding recommendations for fluoroscopy use. Pellegrino, P.L., Brunetti, N.D., Gravina, D., Sacchetta, D., De Sanctis, V., Panigada, S., Di Biase, L., Di Biase, M., and Mantica, M. (2013). Nonfluoroscopic mapping reduces radiation exposure in ablation of atrial fibrillation. *Journal of cardiovascular medicine* 14, 528-533.

Earley, M.J., Showkathali, R., Alzetani, M., Kistler, P.M., Gupta, D., Abrams, D.J., Horrocks, J.A., Harris, S.J., Sporton, S.C., and Schilling, R.J. (2006). Radiofrequency ablation of arrhythmias guided by non-fluoroscopic catheter location: a prospective randomized trial. *Eur Heart J* 27, 1223-1229

Third party trademarks used herein are trademarks of their respective owners.

Caution: US law restricts this device to sale by or on the order of a physician.

Important information: Prior to use, refer to the instructions for use supplied with this device for indications, contraindications, side effects, warnings and precautions.

If you no longer wish to receive emails from Biosense Webster, please [click here](#).